

Highly Accurate, Reliable, Stable, Quantitative, and Remotely Controllable Gas Monitoring System

INNOVA 1512

- Selectively measures a wide range of gases/vapors
- Linear response over a wide dynamic range
- Stable and Reliable: ensuring a maximum of only two calibrations a year
- User-friendly: easy calibration, configuration, and viewing/analyzing of data via PC
- Accurate: compensates for temperature and pressure fluctuations, water vapor interference, and interference from other known gases
- Extremely low-volume flushing possible
- Operates immediately: virtually no warm-up time necessary
- Remote control capability via TCP/IP network interface protocol
- Expandable up to 24 locations with INNOVA 1409 Multipoint Sampler: the Gas Monitor can operate as the system controller for full standalone operation



The Photoacoustic Gas Monitor INNOVA 1512 is a highly accurate, reliable, and stable quantitative gas monitoring system. Its measurement system, based on the photoacoustic infrared detection method, is capable of measuring almost any gas that absorbs infrared light.

Gas selectivity is achieved through the use of optical filters. By installing up to five filters, the 1512 can measure the concentration of up to five component gases and water vapor in any air sample. The detection limit is gas-dependent, but is typically in the ppb region. Accuracy of these measurements is ensured by the 1512's ability to compensate for temperature and pressure fluctuations, water vapor interference, and interference from other gases known to be present. Reliability of measurement results can be ensured by regular self tests. This measurement system requires no consumables and very little regular maintenance. For

most applications, recalibration is only necessary one to two times a year.

The monitoring system is easily operated through either the front panel, with its push-buttons and display providing short explanatory texts, or through the PC software. Both interfaces allow the user to configure the monitor, start a measurement sequence, and view the resulting concentration values of specific gases.

The monitor is equipped with standard interfaces: USB, Ethernet, and RS232. These enable the monitor to be integrated into automated process systems.

The 1512 has a built-in pump system that allows samples to be drawn from up to 50 meters away.

Application areas:

- Occupational Health and Safety measurements – of possible production or accumulation of toxic/ carcinogenic substances in working areas
- Monitoring of anesthetic agents in hospitals
- Emission monitoring of greenhouse gases from agricultural production
- Emission monitoring of exhaust from chemical processes
- Indoor Air Quality (IAQ) measurements
- Ventilation and air exchange using tracer gas

Selectivity

The gas selectivity of the 1512 is determined by the optical filters installed in its filter wheel. Because water is nearly always present in ambient air and absorbs infrared light at most wavelengths, it contributes to the total acoustic signal in the analysis cell. Therefore, the monitor is permanently fitted with a special filter that measures water vapor and enables the 1512 to compensate for water vapor interference. By selecting different filters, this technique can also be used to cross-compensate for known interferent gases.

Calibration


After the relevant optical filters are installed, the monitor must be calibrated. This is achieved through easy-to-use menu driven instructions. Thanks to its high stability, calibration of the 1512 is seldom necessary more than once a year. Calibration is performed using either the Calibration Software BZ7002 or directly from the 1512 front panel.

Operation

The 1512 monitoring system is easy to operate using either the application software LumaSoft™ Gas 7820 or 7880 or by using the front panel push-keys (which can be locked and accessed at three levels using passwords). The monitor can be operated as both an online and offline instrument (i.e. standalone operation). Using these user-interfaces with their logical division of information, everything that needs to be defined is achieved prior to starting the monitoring task.

Configuring the Monitor

The Set-up option enables all the parameters necessary to complete the monitoring task to be defined. This includes setting the Sample Integration Times (S.I.T.) option, which enables measurement results to be weighted - sensitivity against speed. When used as a system controller for multipoint monitoring, the same menu enables the setup of the INNOVA 1409's multipoint sampling tasks.



Measurement Cycle

1. The pump draws air from the sampling point through the air filter to flush out the "old" air in the measurement system and replace it with a "new" air sample. The pressure sensor is used to check that the pump sequence is elapsed successfully and to measure the actual air pressure.
2. The "new" air sample is hermetically sealed in the analyses cell by closing the inlet and outlet valves.
3. Light from an infrared light source is reflected off a mirror, passed through a mechanical chopper, which pulsates it, and then through one of the optical filters in the filter wheel.

4. The gas being monitored selectively absorbs the light transmitted by the optical filter. Because the light is pulsating, the gas temperature increases and decreases, causing an equivalent increase and decrease in the pressure of the gas (an acoustic signal) in the closed cell.
5. Two microphones mounted in the cell wall measure this acoustic signal, which is directly proportional to the concentration of the monitored gas present in the cell.
6. The filter wheel turns so that light is transmitted through the next optical filter, and the new signal is measured. The number of times this step is repeated is dependent on the number of gases being measured.
7. The response time is approximately 13 seconds for one gas or water vapor, or approximately 26 seconds if five gases and water vapor are measured.

Starting Measurements

Once the set-up parameters have been defined, measurements can be started immediately or later using a delayed start time. Once started, the monitoring task continues until it is stopped either manually or by using a defined stop time.

Alarms

Two Alarm trigger levels, which provide high alarm limits for each measured gas, can be defined. These can also be linked to audible alarms using the relay outputs. In addition, the application software LumaSoft™ Gas 7820 or 7880 allows four alarm levels to be displayed.

Online Measurement Results

Using one or more of the monitor's standard interfaces, measurement results are transferred directly to a PC. Here they can be displayed on screen as real-time values in tables and graphs (see Fig. 1) or integrated into the process system.

In the 7820/7880 software, the graphs can be configured to display only the desired gases, defined concentration ranges, and results from statistical analyses. Also, when using the 7880 software, all measurement data is stored in user-defined SQL Server 2014 database.

Offline Measurement Results

Gas measurement result data is displayed on the 1512's screen (display memory) as soon as it is available, and is constantly updated. During a task, the 1512 performs running statistical analyses of the measured gas concentrations, calculating a variety of values for each monitored gas.

This data (in Display Memory) can be copied to the Background Memory, which is a non-volatile storage area. The internal memory stores the measurement readings on a gas per gas basis, but also across the sampling channels when applicable.

Data stored in Background Memory can be recalled to Display Memory. From this memory, data can be uploaded to the BZ7003 Offline Software in either excel or text file format or alternatively printed out on a standard printer.

Reliability

Reliability can be ensured by a series of self tests performed by the monitor. The self tests check software, data integrity, and the 1512's components to ensure that they function properly. If a fault is found, it is reported in the measurement results, so that the integrity of the results can be ensured.

If the power supply fails, the 1512 will automatically start up again when power is restored. Measurement data stored in the monitor's memory is not affected by power loss.




Fig. 1: The graphical window shows up to seven graphs. The user selects the data plotted, the scaling, and the style and color of the lines and background to build the graphical window.

Maintenance

The only maintenance tasks necessary are calibration and replacement of the air filter. Both tasks are easily performed. The frequency for changing the air filter depends on the individual applications.

Multiple Point Monitoring

The INNOVA 1512 can be integrated with the INNOVA 1409 Multipoint Sampler to form a monitoring system expandable to up to 24 channels sampled sequentially. The user can decide upon a full standalone operation (the gas monitor is the system controller) or a remote controlled operation from a PC with the LumaSoft 7880 for online monitoring.

Remote Control Option

LumaSense Technologies offers remote control capability through the user's local area network using the LumaSoft™ Gas Single Point 7820 or Multi Point 7880 software. Online access to the measurement data is available via a built in OPC server (alternatively via Microsoft Excel).

Optional Analog/Relay Interface Module UA1374

The functions of the 1512 can be expanded through the additional Analog/Relay Module UA1374.

For each gas, barometric pressure and chamber temperature, the following outputs are available:

- 0 ... 20 mA, 4 ... 20 mA
- 0 ... 10 V (0 ... 5 V with loss of dynamic range)

Accuracy:	Zero Drift: $\pm 0.25\%$
Voltage Output:	$\pm 1.5\%$ of full scale
Current Output:	$\pm 0.5\%$ of full scale
Resolution:	16 bit (0 ... 20 mA and 0 ... 10 V)
Measurement Range:	Range and zero-point are scalable in the software. Maximum load resistance on current output is 800 Ω . Minimum load resistance for the voltage output is 1000 Ω .

The analog outputs are galvanically isolated from the rest of the analyser, but NOT from each other.

With the Analog/Relay Interface Module, 12 alarm relays can be configured: either as two alarm levels for each gas (plus water) on any active sampling channel, or as alarm relays for selective channels on any monitored gas. Furthermore, two alarm relays are available for warning/error messages and for system watchdog function. Max 25 V DC, Max 100 mA.

Ordering Information

Photoacoustic Gas Monitor - Innova 1512

Optical filters necessary for the user's monitoring task can be ordered together with the 1512, and installed by LumaSense Technologies. The 1512 is then delivered zero-point and humidity interference calibrated.

Includes the Following Accessories

AT2177	4m PTFE tubing
DS0759	Particle filter
VF0102	Fuse
BR6022	Set-up tree
Mains Cable	
AS0001	USB cable
BZ7002	Calibration Software
BZ7003	Offline Software
7820	LumaSoft Gas Single Point monitoring software

Instruction Manual (USB flash drive)

Optional Accessories

The 1512 can be span-calibrated for certain gases – contact your local LumaSense Technologies representative for details of the gases for which this can be done.

27 Optical Filters UA0968 – UA0989

UA0936

UA6008

UA6009

UA6010

UA6016

Calibrations

UA0181 Automated Calibration

UA0182 Complex Calibration

UA0183 Advanced Calibration

Multiple Point Monitoring

1403 Multipoint Sampler and Doser

7650 Basic Ventilation Software (included with the 1403)

7651 Advanced Ventilation Software

7880 LumaSoft Gas Multi Point

1409 Multipoint Sampler

Cables, Adapters, and Tubings

WL0950-003 RS232 Interface cable (9pin-9pin) null modem

JP0600 6-pin DIN plug (male) with locking collar for alarm relay

AF0614 PTFE tubing

UA1365 Genie Membrane separator (inline)

UA1374 Analog/Relay Interface Module

JZ0102 37-pin Sub-d to 37-pin screw terminal (for analog relay)

AO1431 I/O cable one meter (for analog relay)

AO1432 I/O cable three meters (for analog relay)

Technical Specifications

Measurement Technique

Photoacoustic infrared spectroscopy.

Your LumaSense sales representative will assist in the selection of suitable optical filters. Details are provided in the Gas Detection Limits chart.

Response Time

Is dependent on the Sample Integration Time (S.I.T.) and the flushing time defined. Please see the examples below:

Measurement Specifications¹

Monitor-Setup	Response Times
S.I.T.: "Normal" (5 s) Flushing: Auto, (tube 1 m)	One gas: ~27 s 5 gases + water: ~60 s
S.I.T.: "Low Noise" (20 s) Flushing: Auto, (tube 1 m)	5 gases + water: ~150 s
S.I.T.: "Fast" (1 s) Flushing: Chamber 4 s, Tube "OFF"	One gas: ~13 s 5 gases + water: ~26 s

Detection Limit: Gas-dependent, but typically in the ppb region. Using the Gas Detection Limits chart, the detection limit for a selected sample integration time (S.I.T.) can be calculated.

Dynamic Range: Typically 4 orders of magnitude (i.e. 10,000 times the detection limit at 5 S.I.T.). Using two span concentrations it can be expanded to 5 orders of magnitude.

Zero Drift: Typically \pm Detection limit⁴ per 3 months¹.

Influence of temperature²: \pm 10% of detection limit⁴/°C.

Influence of pressure³: \pm 0.5% of detection limit⁴/mbar.

Repeatability: 1% of measured value¹

Range Drift: \pm 2.5% of measured value per 3 months¹.

Influence of temperature²: \pm 0.3% of measured value/°C.

Influence of pressure³: - 0.01% of measured value/mbar.

Reference conditions:

¹ Measured at 20 °C, 1013 mbar, and relative humidity (RH): 60%. (A concentration of 100x detection limit⁴ was used in determining these specifications.)

² Measured at 1013 mbar, and RH: 60%.

³ Measured at 20 °C and RH: 60%.

⁴ Detection limit is @ 5 s S.I.T.

Interference:

The 1512 automatically compensates for temperature and pressure fluctuations in

its analysis cell and can compensate for water vapor in the air sample. If an optical filter is installed to measure a known interferent, the 1512 can cross compensate for the interferent.

Acoustic Sensitivity: not influenced by external sound.

Vibration Sensitivity: strong vibrations at 20 Hz can affect the detection limit.

Internal Data Storage Capacity

The total space available in Display Memory to store data is 131072 measurement cycles. If a measurement cycle takes 15 sec, then the display Memory space will be sufficient for a 22-day monitoring task.

General

Pumping Rate: 30 cm³/s (flushing sampling tube) and 5 cm³/s (flushing measurement chamber).

Power ratings: 85 VA, 100 ... 240 V AC \pm 10%, 50 & 60 Hz, Class 1.

Air Volume per sample:

Flushing Settings	Volume of Air
Auto: Tube Length: 1 m	140 cm ³ /sample
Fixed time: Chamber 2 s, Tube 3 s	100 cm ³ /sample
Fixed time: Chamber 2 s, Tube "OFF"	10 cm ³ /sample

Total Internal Volume: The total Internal Volume of the measurement system: 60 cm³

Alarm Relay Socket: for connection to one or two alarm relays (visual/audio). Alarm levels for each gas are user-defined. System On/Running status available. Max. 25 V DC, max.100 mA.

Back-up Battery: 3 V lithium battery, lifetime 5 years. This protects data stored in memory, and powers the internal clock.

Dimensions:

Height: 195 mm (7.68 in)

Width: 443 mm (17.44 in)

Depth: 244 mm (9.60 in)

Weight: 11 kg (24.3 lbs)

Communication

The monitor uses three interfaces, USB, Ethernet, and RS232, for data exchange and remote control of the 1512. The software communicates using the USB, Ethernet, and RS232 interface.

Computer Requirements



Hardware:

Intel dual-core i3 or compatible. Min. 4096 MB RAM. Min. 500 MB space available on hard-disk.

Software:

7820/7880/BZ7002/BZ7003/7650/7651: Windows[®] 7, 8.1, and Windows[®] 10.

WARNING: The 1512 must not be placed in areas with flammable gases/vapors in explosive concentrations or be used to monitor explosive concentrations of these. Monitoring of certain aggressive gases or a very high concentration of water vapor may damage the 1512. Contact your LumaSense sales representative for further information.

 	COMPLIANCE WITH STANDARDS: CE-mark indicates compliance with: EMC Directive and Low Voltage Directive. NEMKO mark indicates compliance with: CSA and UL Standards.	
Safety	EN/IEC 61010-1 3rd Edition CAN/CSA C22.2 No. 61010-1-04 UL 61010-1 3rd Edition	Safety Requirements for electrical equipment for measurement, control, and laboratory use. Safety Requirements for electrical equipment for measurement, control, and laboratory use. Safety Requirements for electrical equipment for measurement, control, and laboratory use.
EMC	EN 61326-1:2013	Electrical equipment for measurement, control and laboratory use – EMC requirements; Part 1: General requirements
Environment	UL 61010A-1: Environmental conditions. Altitude up to 2000 m Operating Temperature: + 5 °C ... + 40 °C Storage Temperature: - 25 °C ... + 55 °C Humidity: Maximum relative humidity 80% for temperatures up to 31 °C decreasing linearly to 50% relative humidity @ 40 °C Pollution Degree 2 Overvoltage Category II Indoor Use	
Enclosure	IP20	

LumaSense Technologies | An Advanced Energy Company

Temperature and Gas Sensing Solutions

Americas, Australia, India, Asia
Sales & Service
Santa Clara, CA
Tel: +1 800 631 0176
Fax: +1 408 727 1677

Europe, Middle East, Africa
Sales & Service
Frankfurt, Germany
Tel: +49 69 97373 0
Fax: +49 69 97373 1677

INNOVA Gas Products Sales & Service
LumaSense Technologies A/S
Energivej 30, DK- 2750 Ballerup
Tel: +45 44 20 01 00
Fax: +45 44 20 01 01

info@lumasenseinc.com

LumaSense Technologies, Inc., reserves the right to change the information in this publication at any time.

www.lumasenseinc.com
©2018 LumaSense Technologies - Innova-1512_Datasheet-EN - Rev. 09/26/2018
All rights reserved. LumaSense Technologies, Inc., a subsidiary of Advanced Energy Industries, Inc.